WARNING

1. Do not use sling if a patient suffers from back injury.
2. Please understand hoist operating instructions before using.
3. Check sling before use for signs of wear.

General Purpose With Head Support Lifting Procedures

KH407N1—20-40kg
KH407N2—40-150kg
KH407N3—150-250kg

STEP 1: When the patient is lying in bed, roll them onto their side and place the sling under them similar to fitting a draw sheet, ie: roll the patient onto their side concertina the sling under the patients back. Align (A) with the base of the patients spine, then roll the patient onto the sling and centralise. If the patient is sitting up in bed or in a chair, lean the patient forward and slide the sling down the back of the bed/ chair. Try to tuck (A) as low as possible under the patient bottom.

STEP 2: Bring the lifter into position, ensure hoist legs are gated and the hoist yoke is approximately over the centre of the patient.
DO NOT engage the brakes on the castors

STEP 3: Attach straps (B) & (C) to yoke hooks. Pass (D) & (E) under the thighs and cross over straps and attach to the opposite side hook. Always use the same colour strap on either side. Please use the record chart on back of manual once the ideal strap position is determined.

Patients can be positioned 2 ways.

1. Sitting upright—(B) & (C) use short straps, (D) & (E) use long straps
2. Reclined lift—(B) & (C) use long straps, (D) & (E) use short straps

Sitting upright lift.
Used to place patient in a chair.

Reclined lift.
Used to place patient in a bed

Note: 1.The operator/carer can steady the patient by holding the back handles.
Note: 2. To prevent patient discomfort ensure leg/thigh sections are not creased/folded under the patients thighs prior to lifting

STEP 4. Start with the lifting process using the slow speed. When lowering a patient into a chair pull on the back strap as necessary to locate the patients back into the chair.
CAUTION: Hoist may move towards the patient when boom is raised.
Position the patient over chair, bed or toilet and begin lowering. Make sure the patient feels comfortable and secure in the sling.

CARE INSTRUCTIONS

1. Sling Care

Protect from direct sunlight.
Do not puncture with any sharp objects.
Repair any tear immediately.
Do not lift a patient if there are any sling defects.
Washing and drying temperature must not exceed 60 degrees.

2. Spot Cleaning

The sling is machine washable.
Treat spills and stains as soon as possible.
Gently scrape away any soils or mop away liquid from surface of fabric with damp cloth.
Do not scrub with stiff brush.
Dry in shade, away from direct heat.
Dry thoroughly before use.

Kerry Equipment™
A Hills Company

4 Halley Rd
Balcatta WA 6021
Tel: 08 9240 2944
Fax: 08 9240 1469
Email: info@kerryequipment.com.au
Website: www.kerryequipment.com.au

PATIENT RECORDS FOR MOUNTING POINTS
USING KH407N1, KH407N2, KH407N3

Name of Patient	Sling size	Laying	Seated	Strap colour (A) & (B)	Strap colour (C) & (D)

USERS GUIDE
KH407N
GENERAL PURPOSE
WITH HEAD SUPPORT

FEATURES
Mesh/netting option KH407NE.
Easy to fit.
Excellent back and head support.
Padding fitted to leg area.
Commode aperture.
Colour coded attachment straps.
Positioning straps.
Machine washable.

Kerry Equipment™
A Hills Company